



ESNORE & SLEEP LLC
 5975 SHILOH RD, STE 114
 ALPHARETTA
 Georgia, 30005

Phone: 4042207635
 Email: rob.brant@esnoreandsleep.com



07/01/2015

DOB: 06/08/1964

Age: 51 years

AirView™

Detailed Report

AirSense 10 AutoSet

Serial number



Line 1 Mode / min - max pressure **AutoSet / 5.0 - 14.0 cmH2O**

EPR / level **Fulltime / 3.0 cmH2O**

Statistics

Usage **6 hr 13 min**

Leak - L/min Median: **0.0** 95th %: **2.4**

Line 2 **Pressure - cmH2O** Median: **8.8** 95th %: **12.0**


Events per hour AHI: **0.1** AI: **0.1** HI: **0.0**

Central: **0.0** Obstructive: **0.1** Unknown: **0.0**

Cheyne-Stokes respiration (duration) **0 min (0.0%)**


Leak (L/min)

— Leak
 --- Set threshold



Line 3 **Pressure (cmH2O)**

— Pressure
 --- Set pressure min/max



Events

| >= 10 sec Hypopnea
 Central
 Obstructive
 Unknown

